



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Patrick C Hudson

Location: Ward County Jail

Intoxilyzer Serial No.: 80-004206

Lot Number: 659357

Cylinder Number: 11

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 079 %

Test 2: 0. 080 %

Test 3: 0. 079 %

(ATTACH TEST RECORD)

Patrick C Hudson  
 Chemical Test Operator Signature

10/2/2009  
 Date Cylinder Changed

Deb Kashner

06 Oct 09 OK

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-004206  
 p. 1 of 2  
 FH


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004206  
Location = WARD      8164.12.00 06/09  
10/02/2009      20:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:58
02 Std. Gas	0.079	20:59
03 Room Air	0.000	20:59
04 Std. Gas	0.080	21:00
05 Room Air	0.000	21:00
06 Std. Gas	0.079	21:01
07 Room Air	0.000	21:01

Lot No = 659357  
Cyl No = 11  
Exp Date = 09/03/2011  
County = 51      Oper No. = 049509



Operator Signature  
PATRICK C. HUDSON

Remarks:

Form 106-I8000

80-004206  
P. 2 of 2  
FHM