



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Coutess, Mark</i>		
Location <i>HARV</i>	Intoxilyzer® Serial Number <i>80-002666</i>	
Gas Lot Number <i>16512080A2</i>	Gas Cylinder Number <i>32</i>	Gas Expiration Date <i>8/1/2014</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S)
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. 079 AC
 - B. 0. 079 AC
 - C. 0. 079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Mark Coutess</i>	Date <i>10/2/2013</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>11 Oct 2013</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
10/02/2013 10:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Gas	0.079	10:57
03 Room Air	0.000	10:58
04 Std. Gas	0.079	10:58
05 Room Air	0.000	10:59
06 Std. Gas	0.079	10:59
07 Room Air	0.000	11:00

Lot No = 16512080A2
Cyl No = 32
Exp Date = 08/01/2014
County = 52 Oper No. = 030819


Operator Signature
MARK COUNTESS

Remarks:

Form 106-I8000