



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <i>Tofelard, Jonathan, M</i>		
Location: <i>MoHL</i>		
Intoxilyzer® Serial Number: <i>80-004198</i>		
Gas Lot Number: <i>659358</i>	Gas Cylinder Number: <i>27</i>	Expiration Date: <i>9-5-11</i>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>072</u> %		
Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

*[Signature]* 4564  
 \_\_\_\_\_  
 Chemical Test Operator Signature

*10-4-10*  
 \_\_\_\_\_  
 Date

*Deb Kashner*  
 \_\_\_\_\_  
 Reviewed By

*27 Oct 2010*  
 \_\_\_\_\_  
 Date

*Poor ACA test -  
 this Intox needs to  
 be sent to the factory.  
 Deb Kashner  
 27 Oct 2010*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = MOHL      8164.13.00 06/09  
10/04/2010      15:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:53
02 Std. Gas	0.079	15:54
03 Room Air	0.000	15:54
04 Std. Gas	0.072	15:55
05 Room Air	0.000	15:55
06 Std. Gas	0.081	15:55
07 Room Air	0.000	15:56

Lot No = 659358  
Cyl No = 27  
Exp Date = 09/03/2011  
County = 38      Oper No. = 100915



Operator Signature  
JONATHAN M. TOFTELAND

Remarks:

Form 106-I8000

80-004198  
P. 103 of 103  
JMT