



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Officer Pete Nielson</u>		
Location: <u>West Fargo PD</u>		
Intoxilyzer® Serial Number: <u>80-003059</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>10</u>	Expiration Date: <u>9-3-11</u> ^{12/23/2011} _{OK}
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

10-5-10
Date

Deb Kashner
Reviewed By

11 Oct 2010
Date

80-003059
P. 97 of 98
JAM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = WFPD 8164.13.00 06/09
10/05/2010 13:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:09
02 Std. Gas	0.080	13:09
03 Room Air	0.000	13:09
04 Std. Gas	0.079	13:10
05 Room Air	0.000	13:10
06 Std. Gas	0.081	13:11
07 Room Air	0.000	13:11

Lot No = 659357
Cyl No = 10
Exp Date = 12/23/2011
County = 09 Oper No. = 039617



Operator Signature
PETER NIELSEN

Remarks:

Form 106-18000

80-003059
P. 98-4-98
FM