



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Officer Pete Nielson</u>		
Location: <u>West Fargo PD</u>		
Intoxilyzer® Serial Number: <u>80-003059</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>10</u>	Expiration Date: <u><del>9-3-11</del> <sup>12/23/2011</sup> <del>OK</del></u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	<u>09/03/11</u> <small>OK 29 Apr 11</small>
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

[Signature]  
Chemical Test Operator Signature

10-5-10  
Date

Deb Kashner  
Reviewed By

11 Oct 2010  
Date

80-003059  
P. 108 of 109  
FM  
80-003059  
P. 97 of 98  
FM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = WFPD      8164.13.00 06/09  
10/05/2010      13:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:09
02 Std. Gas	0.080	13:09
03 Room Air	0.000	13:09
04 Std. Gas	0.079	13:10
05 Room Air	0.000	13:10
06 Std. Gas	0.081	13:11
07 Room Air	0.000	13:11

Lot No = 659357  
Cyl No = 10  
Exp Date = ~~12/23/2011~~ *09/03/11* *DK 29 Apr 11*  
County = 09      Oper No. = 039617

  
\_\_\_\_\_  
Operator Signature  
PETER NIELSEN

Remarks:

Form 106-I8000

80-003059  
P.109 of 109  
SM  
80-003059  
P.98 of 98  
SM