



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>DUISEN, SHANNON P</u>		
Location: <u>TIOGA P.P.</u>		
Intoxilyzer® Serial Number: <u>80-004938</u>		<u>DK</u>
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>21</u>	Expiration Date: <u>123111 12/23/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>077</u> %		
Test 2: 0. <u>077</u> %		
Test 3: 0. <u>077</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

10/5/10
Date

Deb Kashur
Reviewed By

06 Oct 2010
Date

80-004938
P. 55 of 56
1/11/11

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TIOG 8164.13.00 06/09
10/05/2010 11:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Gas	0.077	11:01
03 Room Air	0.000	11:01
04 Std. Gas	0.077	11:02
05 Room Air	0.000	11:02
06 Std. Gas	0.077	11:02
07 Room Air	0.000	11:03

Lot No = 709352
Cyl No = 21
Exp Date = 12/23/2011
County = 53 Oper No. = 020632



Operator Signature
SHANNON P. DUISEN

Remarks:

Form 106-I8000

80-004938
P. 56 of 56
JPM