



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Adam Krabbenhoff</u>		
Location: <u>Lamoure county sheriff's office</u>		
Intoxilyzer® Serial Number: <u>80-004942</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>14</u>	Expiration Date: <u>12-23-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>079</u> % Test 2: 0. <u>078</u> % Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

[Signature]
 Chemical Test Operator Signature

10-5-10
 Date

Deb Kashur
 Reviewed By

11 Oct 2010
 Date


Intoxilyzer Test Record and Checklist
NDODAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = LAMO 8164.13.00 06/09
10/05/2010 14:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:14
02 Std. Gas	0.079	14:14
03 Room Air	0.000	14:15
04 Std. Gas	0.078	14:15
05 Room Air	0.000	14:15
06 Std. Gas	0.079	14:16
07 Room Air	0.000	14:16

Lot No = 709352
Cyl No = 14
Exp Date = 12/23/2011
County = 23
Oper No. = 020028



Operator Signature
ADAM R. KRABBENHOFT

Remarks: external
printer not on.

Form 106-18000

80-004942
P. 52 of 52
jam