



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

|  |                                |                                |
|--|--------------------------------|--------------------------------|
| Chemical Test Operator Name: <u>David M. Walker</u>  |                                |                                |
| Location: <u>Dickinson JPL</u>   |                                |                                |
| Intoxilyzer® Serial Number: <u>80-004947</u>   |                                |                                |
| Gas Lot Number: <u>159358</u>  | Gas Cylinder Number: <u>23</u> | Expiration Date: <u>9-3-11</u> |
| <input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed |                                |                                |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%)  |                                |                                |
| Test 1: <u>0.080</u> %   |                                |                                |
| Test 2: <u>0.081</u> %   |                                |                                |
| Test 3: <u>0.080</u> %   |                                |                                |
| (ATTACH TEST RECORD)   |                                |                                |

David M. Walker  
Chemical Test Operator Signature

10-5-10  
Date

Deb Kashur  
Reviewed By

06 Oct 2010  
Date

80-004947  
P. 51 of 52  
JKM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = DKPD      8164.13.00 06/09  
10/05/2010      07:36

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 07:37 |
| 02 Std. Gas | 0.080 | 07:37 |
| 03 Room Air | 0.000 | 07:38 |
| 04 Std. Gas | 0.081 | 07:38 |
| 05 Room Air | 0.000 | 07:39 |
| 06 Std. Gas | 0.080 | 07:39 |
| 07 Room Air | 0.000 | 07:39 |

Lot No = 659358  
Cyl No = 23  
Exp Date = 09/03/2011  
County = 45      Oper No. = 068332



Operator Signature  
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004947  
P. 52 of 52  
RM