



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Lt. Tracy L. Meidinger</i>		
Location <i>UND-PO</i>	Intoxilyzer® Serial Number <i>80-004937</i>	
Gas Lot Number <i>11611080A2</i>	Gas Cylinder Number <i>138</i>	Gas Expiration Date <i>06/01/2013</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.080* AC
 - B. *0.081* AC
 - C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Lt. Tracy L. Meidinger #258</i>	Date <i>10-05-11</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>03 Apr 12</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = UNDP 8164.13.00 06/09
10/05/2011 22:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:07
02 Std. Gas	0.080	22:07
03 Room Air	0.000	22:08
04 Std. Gas	0.081	22:08
05 Room Air	0.000	22:09
06 Std. Gas	0.081	22:09
07 Room Air	0.000	22:10

Lot No = 11611080A2
Cyl No = 138
Exp Date = 06/01/2013
County = 18 Oper No. = 119409

 #258
Operator Signature
TRACY L. MEIDINGER

Remarks:
changed cylinder
AEA