

INTOXILYZER 8000 INSPECTION

Serial No.: 80-002669

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
 Sim SN DR5132 Lot No. Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
 Sim SN DR3850 Lot No. Rep AC 0.297 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
 (Attach test record.)
 Sim SN DR3846 Lot No. Rep AC 0.10 AC + acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
 (Attach test record.) ✓

K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
 (Attach test record.)

Lot No. <u>15410080A1</u>	Cyl No. <u>25</u>	Exp. Date <u>07/01/2012</u>
Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average _____

L. Tank pressure, Level 3 Func D Sub G P: 218 psi Reg 250 psi

M. Remarks/Maintenance record: Annual inspection

26 Sept 2011
 Date

Deb Keshum
 Field Inspector's Signature

06 Oct. 11
 Date

Charles E. E...
 Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
09/26/2011 09:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 25
Standard Expiration? 07/01/2012
Oper No? 777777

Flow Cal. Date: 09/09/2011
Slope 662
Intercept -510765

IR Calibration Date: 09/09/2011
 3um 9um

0th Coef(*100): -15623 -9015
1st Coef(*100): 263092 133867
2nd Coef(*100): 3130 1285
H2O adj(mg/l*10k): 642 392

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
09/26/2011 09:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:26
02 Std. Sol.	0.019	09:27
03 Room Air	0.000	09:27
04 Std. Sol.	0.020	09:28
05 Room Air	0.000	09:29
06 Std. Sol.	0.019	09:29
07 Room Air	0.000	09:30

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.021AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
09/26/2011 09:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:31
02 Std. Sol.	0.296	09:32
03 Room Air	0.000	09:33
04 Std. Sol.	0.297	09:34
05 Room Air	0.000	09:34
06 Std. Sol.	0.297	09:35
07 Room Air	0.000	09:35

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Keshur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.297 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
09/26/2011 09:37

Test	AC	Time
01 Room Air	0.000	09:38
02 *Subject Test	INT*	09:39
03 Room Air	0.000	09:39

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 15410080A1
Cyl No = 25
Expiration Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10 AC + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
09/26/2011 09:39

Test	AC	Time
01 Diagnostic	OK	09:40
02 Room Air	RFI*	09:40
03 Room Air	0.000	09:41

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 15410080A1
Cyl No = 25
Expiration Date = 07/01/2012
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
09/26/2011 09:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:44
02 Std. Gas	0.080	09:44
03 Room Air	0.000	09:44
04 Std. Gas	0.081	09:45
05 Room Air	0.000	09:45
06 Std. Gas	0.080	09:46
07 Room Air	0.000	09:46

Lot No = 15410080A1
Cyl No = 25
Exp Date = 07/01/2012
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check 0080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
09/26/2011 09:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:47
02 Std. Gas	0.081	09:47
03 Room Air	0.000	09:48
04 Std. Gas	0.081	09:48
05 Room Air	0.000	09:49
06 Std. Gas	0.081	09:49
07 Room Air	0.000	09:49

Lot No = 15410080A1
Cyl No = 25
Exp Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
09/26/2011 09:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:50
02 Std. Gas	0.081	09:51
03 Room Air	0.000	09:51
04 Std. Gas	0.081	09:52
05 Room Air	0.000	09:52
06 Std. Gas	0.081	09:53
07 Room Air	0.000	09:53

Lot No = 15410080A1
Cyl No = 25
Exp Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000