



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Will Vance

Location: DICKINSON LEC

Intoxilyzer Serial No.: 80-004184

Lot Number: 659358

Cylinder Number: 23

Expiration Date: 9/3/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 079 %

Test 2: 0. 080 %

Test 3: 0. 079 %

(ATTACH TEST RECORD)

Will Vance
 Chemical Test Operator Signature

10/7/09
 Date Cylinder Changed

Deb Kashner
 Reviewed By

16 Oct 09
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004184
 p. 1 of 2
 km

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = DKPD 8164.12.00 06/09
10/07/2009 10:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:37
02 Std. Gas	0.079	10:38
03 Room Air	0.000	10:38
04 Std. Gas	0.080	10:39
05 Room Air	0.000	10:39
06 Std. Gas	0.079	10:40
07 Room Air	0.000	10:40

Lot No = 659358
Cyl No = 23
Exp Date = 09/03/2011
County = 45 Oper No. = 019620

William D. Vance

Operator Signature
WILLIAM D. VANCE

Remarks: *ACA*

Form 106-I8000

*80-004184
P. 2 of 2 JAV*