

✓ Database
CML

**INTOXILYZER® 5000 G AND 8000
INSTALLATION AND REPAIR CHECKOUT**

Serial No.: 80-004184 Instrument Location: Dickinson PD

Reason for Install/Repair: Install After Lab Repair/Inspection
 Other (specify) _____

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® KB.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Ethanol Breath Standard: _____ Connected Installed.
(Include Form 105-G if New Cylinder Installed.)
- F. Review/Enter Gas Setup (Level 1, Function S). (Intoxilyzer® 8000 Only.)
- G. Print Test (Level 1, Function P). (Attach Test Record.)
- H. ACA Test (Level 1, Function C). (Attach Test Record.)
- I. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- J. Repair and/or Maintenance Performed: _____

Install new Intox 8000 @ Dickinson LEC

10/7/2009
Date
16 Oct 09
Date

W.D. Vance
Field Inspector's Signature
Deb Kashner
Reviewed By


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = DKPD 8164.12.00 06/09
10/07/2009 10:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:31
02 Std. Gas	0.078	10:31
03 Room Air	0.000	10:32
04 Std. Gas	0.079	10:32
05 Room Air	0.000	10:33
06 Std. Gas	0.080	10:33
07 Room Air	0.000	10:34

Lot No = 659358
Cyl No = 23
Exp Date = 09/03/2011
County = 45 Oper No. = 019620



Operator Signature
WILLIAM D. VANCE

Remarks: *New cylinder*

Form 106-I8000

80-004184
P. 3 of 4 *WLV*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = DKPD 8164.12.00 06/09
10/07/2009 10:43

Test	AC	Time
01 Diagnostic	OK	10:46
02 Room Air	RFI*	10:46
03 Room Air	0.000	10:46

*Invalid Test
Inhibited - RFI

Sub Name = RFI, TEST NA
Sub DOB = 01/01/1990
Sub Sex = Male Weight = NA
Test = DUI Cit = 111111
Dr. Lic. = ND/11111
Lot No = 659358
Cyl No = 23
Expiration Date = 09/03/2011
County = 45 Oper No. = 019620

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
WILLIAM D. VANCE

Remarks: RFI TEST

Form 106-I8000

80-004184
P. 4 of 4