

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004186  
Location = BSPD      8164.12.00 06/09  
10/07/2009      10:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Gas	0.081	11:01
03 Room Air	0.000	11:01
04 Std. Gas	0.081	11:02
05 Room Air	0.000	11:02
06 Std. Gas	0.082	11:02
07 Room Air	0.000	11:03

Lot No = 659358  
Cyl No = 13  
Exp Date = 09/03/2011  
County = 08      Oper No. = 029710

  
\_\_\_\_\_  
Operator Signature  
PATRICK M. RENZ

Remarks:

ACA

Form 106-I8000

80-004186  
P. 1 of 1 FM