



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 58282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Jamey Jahner</u>		
Location: <u>Cass County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004946</u>		
Gas Lot Number: <u>668180</u>	Gas Cylinder Number: <u>76</u>	Expiration Date: <u>09-21-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.078</u> %		
Test 2: <u>0.078</u> %		
Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

Jamey Jahner
 Chemical Test Operator Signature

10-07-10
 Date

Deb Kashner
 Reviewed By

21 Oct 2010
 Date

80-004946
 P. 59460
 JAH


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = CASS 8164.13.00 06/09
10/07/2010 16:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:55
02 Std. Gas	0.078	16:55
03 Room Air	0.000	16:56
04 Std. Gas	0.078	16:56
05 Room Air	0.000	16:57
06 Std. Gas	0.078	16:57
07 Room Air	0.000	16:58

Lot No = 668180
Cyl No = 76
Exp Date = 09/21/2011
County = 09 Oper No. = 100017



Operator Signature
JAMEY JAHNER

Remarks:

Form 106-I8000

80-004946
P. 00460
jm