

Database
CML ✓

**INTOXILYZER® 5000 G AND 8000
INSTALLATION AND REPAIR CHECKOUT**


Serial No.: 80-~~00498~~ <sup>004189 CML
OK</sup> Instrument Location: WATF

Reason for Install/Repair: _____ Install After Lab Repair/Inspection
X Other (specify) Set up @ Sheriff's Office

Check When Done:

- ✓ A. Surge Protector Installed/Properly Grounded.
- ✓ B. Telephone Line Connected to Intoxilyzer® KB.
- X C. Breath Tube Heated.
- X D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- X E. Ethanol Breath Standard: X Connected L Installed.
(Include Form 105-G if New Cylinder Installed.)
- X F. Review/Enter Gas Setup (Level 1, Function S). (Intoxilyzer® 8000 Only.)
- X G. Print Test (Level 1, Function P). (Attach Test Record.)
- X H. ACA Test (Level 1, Function C). (Attach Test Record.)
- X I. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- J. Repair and/or Maintenance Performed: Installation
@ WATF

10-8-09
Date
06 NOV 09
Date


Field Inspector's Signature
Deb Kashur
Reviewed By

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = WATF 8164.12.00 06/09
10/08/2009 13:44

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659357
Standard Cyl #? 44
Standard Expiration? 09/03/2011
Oper No? 068105

Flow Cal. Date: 06/03/2009
Slope 666
Intercept -550210

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100): -13136 -24026
1st Coef(*100): 256035 131167
2nd Coef(*100): 2113 1064
H2O adj(mg/l*10k): 465 385

***** Printer Test End *****


Operator Signature
JOHN L. FULWIDER

Remarks: *Intoxilyzer setup @ WATF*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = WATF 8164.12.00 06/09
10/08/2009 13:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:56
02 Std. Gas	0.080	13:56
03 Room Air	0.000	13:57
04 Std. Gas	0.080	13:57
05 Room Air	0.000	13:58
06 Std. Gas	0.080	13:58
07 Room Air	0.000	13:59

Lot No = 659357
Cyl No = 44
Exp Date = 09/03/2011
County = 27 Oper No. = 068105


Operator Signature
JOHN L. FULWIDER

Remarks:

ACA
Test

Form 106-I8000

80-004189
p. 3 of 4
FIL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = WATF 8164.12.00 06/09
10/08/2009 14:00

Test	AC	Time
01 Diagnostic	OK	14:03
02 Room Air	RFI*	14:03
03 Room Air	RFI*	14:03

*Invalid Test
 Inhibited - RFI

Sub Name = TEST, SUBJECT NA
Sub DOB = 12/04/1955
Sub Sex = Male Weight = 250
Test = OTH Cit = 12345
Dr. Lic. = ND/12345
Lot No = 659357
Cyl No = 44
Expiration Date = 09/03/2011
County = 27 Oper No. = 068105

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JOHN L. FULWIDER

Remarks:

RFI Test

Form 106-I8000

80-004189
P. 4 of 4
FM