



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: RODNEY C. SAGEN

Location: MERCER COUNTY SHERIFF'S DEPT STANTON N.D.

Intoxilyzer Serial No.: 80-004194

Lot Number: 659357

Cylinder Number: #43

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082 %

Test 2: 0.081 %

Test 3: 0.081 %

(ATTACH TEST RECORD)

Rodney C. Sagen  
 Chemical Test Operator Signature

10/08/2009  
 Date Cylinder Changed

Mary Pearson  
 Reviewed By

10/13/09  
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-004194  
 P. 1 of 2 FR

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004194  
Location = STAT      8164.12.00 06/09  
10/08/2009      15:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:34
02 Std. Gas	0.082	15:34
03 Room Air	0.000	15:35
04 Std. Gas	0.081	15:35
05 Room Air	0.000	15:36
06 Std. Gas	0.081	15:36
07 Room Air	0.000	15:37

Lot No = 659357

Cyl No = 43

Exp Date = 09/03/2011

County = 29

Oper No. = 068810



Operator Signature  
RODNEY C. SAGEN

Remarks:

Form 106-I8000

80-004194  
P. J. J. F. M.