



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Richard D. Johnson

Location: McLean County Sheriff's Dept

Intoxilyzer Serial No.: 80-004197

Lot Number: 659357

Cylinder Number: 47

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

DK 0.081
 Test 1: ~~0.081~~ %

0.082
 Test 2: ~~0.082~~ %

0.081
 Test 3: ~~0.081~~ %

(ATTACH TEST RECORD)

Richard D. Johnson
 Chemical Test Operator Signature

10/08/2009
 Date Cylinder Changed

Deb Kestler
 Reviewed By

14 Oct 09
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004197
 p 1 of 2 PK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004197
Location = WASH 8164.12.00 06/09
10/08/2009 16:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:19
02 Std. Gas	0.081	16:19
03 Room Air	0.000	16:20
04 Std. Gas	0.082	16:20
05 Room Air	0.000	16:21
06 Std. Gas	0.081	16:21
07 Room Air	0.000	16:21


Lot No = 659357

Cyl No = 47

Exp Date = 09/03/2011

County = 28

Oper No. = 118707


Operator Signature
RICHARD D. JOHNSON

Remarks:

Form 106-18000

80-004197
P. 20/2 KIL