



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Josh Rude

Location: Stutsman Co.

Intoxilyzer Serial No.: 80-004203

Lot Number: 659358

Cylinder Number: 48

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082%

Test 2: 0.083%

Test 3: 0.082%

(ATTACH TEST RECORD)

Josh Rude  
 Chemical Test Operator Signature

10-08-09  
 Date Cylinder Changed

Deb Kashner  
 Reviewed By

26 Oct 2009  
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-004203  
 P.1 of 2 JRM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = JAME      8164.12.00 06/09  
10/08/2009      16:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:32
02 Std. Gas	0.082	16:32
03 Room Air	0.000	16:33
04 Std. Gas	0.083	16:33
05 Room Air	0.000	16:33
06 Std. Gas	0.082	16:34
07 Room Air	0.000	16:34

Lot No = 659358  
Cyl No = 48  
Exp Date = 09/03/2011  
County = 47      Oper No. = 100414

  
Operator Signature  
JOSHUA J. RUDE

Remarks:

Form 106-I8000

80-004203  
P. 2 of 2 ym