



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Benny Vannoth

Location: Muhl

Intoxilyzer Serial No.: 80-004204

Lot Number: 659358

Cylinder Number: 8

Expiration Date: 09-03-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 083 %

Test 2: 0. 083 %

Test 3: 0. 083 %

(ATTACH TEST RECORD)

Benny Vannoth  
 Chemical Test Operator Signature

10-8-09  
 Date Cylinder Changed

Deb Kashur  
 Reviewed By

13 Oct 09  
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-004204  
 P. 1 of 2 FK


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = MOHL      8164.12.00 06/09  
10/08/2009      13:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:03
02 Std. Gas	0.083	13:03
03 Room Air	0.000	13:04
04 Std. Gas	0.083	13:04
05 Room Air	0.000	13:04
06 Std. Gas	0.083	13:05
07 Room Air	0.000	13:05

Lot No = 659358  
Cyl No = 8  
Exp Date = 09/03/2011  
County = 38      Oper No. = 039909

  
Operator Signature  
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004204  
P. 2 of 2 JH