



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Phil Wolf</u>		
Location: <u>Cavalier, ND</u>		
Intoxilyzer® Serial Number: <u>80-002669</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>4</u>	Expiration Date: <u>9/3/11</u>
<input type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.076</u> %		
Test 2: <u>0.076</u> %		
Test 3: <u>0.076</u> %		
(ATTACH TEST RECORD)		

P. Wolf
Chemical Test Operator Signature

10/8/10
Date

Deb Kashur
Reviewed By

18 Oct 2010
Date

80-002669
P. 98 of 99
jm


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CAVA 8164.13.00 06/09
10/08/2010 17:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:09
02 Std. Gas	0.076	17:09
03 Room Air	0.000	17:10
04 Std. Gas	0.076	17:10
05 Room Air	0.000	17:11
06 Std. Gas	0.076	17:11
07 Room Air	0.000	17:12

Lot No = 659357
Cyl No = 9
Exp Date = 09/03/2011
County = 34 Oper No. = 120547



Operator Signature
PHIL WOLF

Remarks:

Form 106-I8000

80-002669
R 99499
EM