



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <u>Rick W Hinton</u> <u>Cass Jail - 09</u>		
Location <u>Cass Jail - 09</u>	Intoxilyzer® Serial Number <u>80-004204</u> <u>DK</u>	
Gas Lot Number <u>23411080A1</u>	Gas Cylinder Number <u>71</u>	Gas Expiration Date <u>10-1-13</u>

Check When Done:

1. Scan/Enter Gas Cylinder Information (Level 1, Function S).

2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. 0.080 AC

B. 0.080 AC

C. 0.080 AC

3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

5. Send the Following to the Crime Laboratory:

A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).

B. ACA Test Record.

Chemical Test Operator Signature <u>[Signature]</u>	Date <u>10-8-12</u>
Reviewed By (Crime Laboratory Use Only) <u>Deb Kasher</u>	Date <u>21 NOV 2012</u>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CASS 8164.13.00 06/09
10/08/2012 10:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:06
02 Std. Gas	0.080	10:07
03 Room Air	0.000	10:07
04 Std. Gas	0.080	10:08
05 Room Air	0.000	10:08
06 Std. Gas	0.080	10:08
07 Room Air	0.000	10:09

Lot No = 23411080A1
Cyl No = 71
Exp Date = 10/01/2013
County = 09 Oper No. = 029920



Operator Signature
RICK HINTON

Remarks: