

INTOXILYZER® 5000 G AND 8000
INSTALLATION AND REPAIR CHECKOUT

Serial No.: 80-004193 Instrument Location: GRAF

Reason for Install/Repair: _____ Install After Lab Repair/Inspection
X Other (specify) Re-install after removed from

facility to give demonstration to Walsh Co. State's Atty

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® KB. (NA) not available at install
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Ethanol Breath Standard: X Connected X Installed. (Include Form 105-G if New Cylinder Installed.)
- F. Review/Enter Gas Setup (Level 1, Function S). (Intoxilyzer® 8000 Only.)
- G. Print Test (Level 1, Function P). (Attach Test Record.)
- H. ACA Test (Level 1, Function C). (Attach Test Record.)
- I. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- J. Repair and/or Maintenance Performed: _____

10/09/09
Date
16 Oct 09
Date

Sgt. J. Aldenburg 274 NDA#2
Field Inspector's Signature
Deb Kashner
Reviewed By

80-004193
P.1-14 PL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = GRAF 8164.12.00 06/09
10/09/2009 14:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:48
02 Std. Gas	0.081	14:48
03 Room Air	0.000	14:49
04 Std. Gas	0.080	14:49
05 Room Air	0.000	14:50
06 Std. Gas	0.081	14:50
07 Room Air	0.000	14:51

Lot No = 659358

Cyl No = 16

Exp Date = 09/03/2011

County = 50

Oper No. = 048915

Sgt. D. Oldenburg 274 NDHP

Operator Signature
DOLF H. OLDENBURG

Remarks:

Form 106-I8000

80-004193
p. 3-14 PM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = GRAF 8164.12.00 06/09
10/09/2009 14:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:58
02 Std. Gas	RFI*	14:58
03 Room Air	0.000	14:58

*Invalid Test
Inhibited - RFI

Lot No = 659358

Cyl No = 16

Exp Date = 09/03/2011

County = 50

Oper No. = 048915

Sgt. D. Oldenburg 274 NDHP

Operator Signature
DOLF H. OLDENBURG

Remarks:

Form 106-I8000

80-004193
R 4/4 FM