



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Melby, Robert M.

Location: CROS (Divide Co. S.O.)

Intoxilyzer Serial No.: 80-002667

Lot Number: 659358

Cylinder Number: 6

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

DFD. 079  
 Test 1: 0. 79 %  
 0.079  
 Test 2: 0. 79 %  
 0.079  
 Test 3: 0. 79 %

(ATTACH TEST RECORD)

R. Melby  
 Chemical Test Operator Signature

10/09/09  
 Date Cylinder Changed

Deb Kashur  
 Reviewed By

16 Oct 09  
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

*P. Tafel*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = CROS      8164.12.00 06/09  
10/09/2009      11:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:31
02 Std. Gas	0.079	11:31
03 Room Air	0.000	11:32
04 Std. Gas	0.079	11:32
05 Room Air	0.000	11:33
06 Std. Gas	0.079	11:33
07 Room Air	0.000	11:33

Lot No = 659358  
Cyl No = 6  
Exp Date = 09/03/2011  
County = 12      Oper No. = 100011


  
\_\_\_\_\_  
Operator Signature  
ROBERT M. MELBY

Remarks:

Test OK

Form 106-I8000

The undersigned, having custody of the original record, certifies that the information hereon is a true and correct copy of the original document maintained as part of the activities of this office.

      10/09/09  
Signature      Date

P. 2 of 2 km