

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CASS 8164.12.00 06/09
10/09/2009 01:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:55
02 Std. Gas	0.082	01:55
03 Room Air	0.000	01:56
04 Std. Gas	0.081	01:56
05 Room Air	0.000	01:56
06 Std. Gas	0.082	01:57
07 Room Air	0.000	01:57

Lot No = 659357
Cyl No = 40
Exp Date = 09/03/2011
County = 09 Oper No. = 020535

S Getz

Operator Signature
SHAWN P. GETZ

Remarks: *ACA after reboot ... froze up
on 3513*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CASS 8164.12.00 06/09
10/09/2009 01:46

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659357
Standard Cyl #? 40
Standard Expiration? 09/03/2011
Oper No? 020535

Flow Cal. Date: 09/10/2007
Slope 702
Intercept -608746

IR Calibration Date: 09/10/2007
 3um 9um

0th Coef(*100): -4604 -8890
1st Coef(*100): 254947 133181
2nd Coef(*100): 4056 1165
H2O adj(mg/l*10k): 616 347

***** Printer Test End *****

Shawn P. Getz

Operator Signature
SHAWN P. GETZ

Remarks: *Print test after re-boot*