



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Eugene M Moheit

Location: H & TT Adams Co SD

Intoxilyzer Serial No.: 80-003072

Lot Number: 1659358

Cylinder Number: 7

Expiration Date: 09-03-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 083 %

Test 2: 0. 083 %

Test 3: 0. 082 %

(ATTACH TEST RECORD)

Eugene M Moheit  
 Chemical Test Operator Signature

10-9-2009  
 Date Cylinder Changed

Mary Pearson  
 Reviewed By

10-13-2009  
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-003072  
 p. 1 of 2 pm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = HETT      8164.12.00 06/09  
10/09/2009      09:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:55
02 Std. Gas	0.083	09:55
03 Room Air	0.000	09:56
04 Std. Gas	0.083	09:56
05 Room Air	0.000	09:57
06 Std. Gas	0.082	09:57
07 Room Air	0.000	09:57


Lot No = 659358

Cyl No = 7

Exp Date = 09/03/2011

County = 01

Oper No. = 049130

  
Operator Signature  
EUGENE M. MOLBERT

Remarks:

Form 106-I8000

80-003072  
P. 2 of 2 FM