



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (5-2011)

Serial Number 80-005947	Instrument Location Richland County Jail
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (Use CMS Mode; Key Radio During First Room Air).
- 7. Repair and/or Maintenance Performed (if Any): NONE
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date 10/02/2013 10/11/2013
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashur</i>	Date 14 Oct 2013

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = WAHP 8164.13.00 06/09
10/02/2013 13:59

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 58
Standard Expiration? 08/01/2014
Oper No? 020401

Flow Cal. Date: 06/21/2013
Slope 693
Intercept -735360

IR Calibration Date: 06/22/2013
 3um 9um

0th Coef(*100):	-23353	-21151
1st Coef(*100):	285912	140248
2nd Coef(*100):	2576	1232
H2O adj(mg/l*10k):	602	508

***** Printer Test End *****



Operator Signature
MATTHEW ANDERSON

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = WAHP 8164.13.00 06/09
10/02/2013 14:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:00
02 Std. Gas	0.082	14:01
03 Room Air	0.000	14:01
04 Std. Gas	0.082	14:02
05 Room Air	0.000	14:02
06 Std. Gas	0.081	14:03
07 Room Air	0.000	14:03

Lot No = 16512080A2
Cyl No = 58
Exp Date = 08/01/2014
County = 39 Oper No. = 020401



Operator Signature
MATTHEW ANDERSON

Remarks:

2nd AA test good test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = WAHP 8164.13.00 06/09
10/02/2013 14:04

Test	AC	Time
01 Room Air	RFI*	14:06
02 Room Air	RFI*	14:06

*Invalid Test
 Inhibited - RFI

Sub Name = RFI TEST, RFI TEST NMN
Sub DOB = 10/02/2000
Sub Sex = Male Weight = 120
Test = OTH Cit = 8888888
Dr. Lic. = ND/77777777
Lot No = 16512080A2
Cyl No = 58
Expiration Date = 08/01/2014
County = 39 Oper No. = 020401



Operator Signature
MATTHEW ANDERSON

Remarks:

RFI test, Good test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = WAHP 8164.13.00 06/09
10/11/2013 14:05

Test	AC	Time
01 Diagnostic	OK	14:07
02 Room Air	RFI*	14:07
03 Room Air	RFI*	14:07

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, RFI TEST NMN
Sub DOB = 10/11/1990
Sub Sex = Male Weight = 125
Test = DUI Cit = 11111111
Dr. Lic. = ND/NO DL
Lot No = 16512080A2
Cyl No = 58
Expiration Date = 08/01/2014
County = 39 Oper No. = 020401

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
MATTHEW ANDERSON

Remarks:

RFI test
Form 106-I8000