



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Shannon Liebold</u>		
Location: <u>Monroe County Sheriff's Dept.</u>		
Intoxilyzer® Serial Number: <u>80-003060</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>43</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed		
<input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.081</u> %		
Test 2: <u>0.080</u> %		
Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

Shannon Liebold
Chemical Test Operator Signature

10-12-10
Date

Deb Kashur
Reviewed By

21 Oct 2010
Date

80-003060
P. 93 of 95
IAM



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Shannon Liebold</u>		
Location: <u>Mercer County Sheriff's Dept.</u>		
Intoxilyzer® Serial Number: <u>80-003060</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>43</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	-
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0.____%		
Test 2: 0.____%		
Test 3: 0.____%		
(ATTACH TEST RECORD)		

Shannon P Liebold
Chemical Test Operator Signature

10-12-10
Date

Reviewed By

Date

80-003060
P. 94/95
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = STAT 8164.13.00 06/09
10/12/2010 15:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:29
02 Std. Gas	0.081	15:29
03 Room Air	0.000	15:30
04 Std. Gas	0.080	15:30
05 Room Air	0.000	15:31
06 Std. Gas	0.080	15:31
07 Room Air	0.000	15:31

Lot No = 659357
Cyl No = 43
Exp Date = 09/03/2011
County = 29 Oper No. = 020904



Operator Signature
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-003060
P. 95 of 95
PAN