



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Hinton, Rick W</u>		
Location: <u>Cass County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004946</u>		
Gas Lot Number: <u>668180</u>	Gas Cylinder Number: <u>76</u>	Expiration Date: <u>9-21-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.078</u> %		
Test 2: <u>0.077</u> %		
Test 3: <u>0.076</u> %		
(ATTACH TEST RECORD)		

Rick W Hinton
Chemical Test Operator Signature

10/12/10
Date

Deb Kashur
Reviewed By

22 Oct 2010
Date

80-00 4946
P. 614/62
RM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = CASS 8164.13.00 06/09
10/12/2010 08:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:06
02 Std. Gas	0.078	08:06
03 Room Air	0.000	08:07
04 Std. Gas	0.077	08:07
05 Room Air	0.000	08:07
06 Std. Gas	0.076	08:08
07 Room Air	0.000	08:08

Lot No = 668180
Cyl No = 76
Exp Date = 09/21/2011
County = 09 Oper No. = 029920



Operator Signature
RICK HINTON

Remarks:

Form 106-I8000

80-004946
P. 62 of 62
JEM