



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	William R. Petty				
Location:	GF AF				
Intoxilyzer® Serial Number:	80-002667				
Gas Lot Number:	65 9358	Gas Cylinder Number:	39	Expiration Date:	09/03/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.08 %					
Test 2: 0.08 %					
Test 3: 0.08 %					
(ATTACH TEST RECORD)					

Will Petty
Chemical Test Operator Signature

13 Oct 10
Date

Deb Kashur
Reviewed By

27 Oct 2010
Date

80-002667
P. 7 of 79
FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = GFAP 8164.13.00 06/09
10/13/2010 10:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Gas	0.080	10:36
03 Room Air	0.000	10:36
04 Std. Gas	0.080	10:37
05 Room Air	0.000	10:37
06 Std. Gas	0.080	10:38
07 Room Air	0.000	10:38

Lot No = 659358
Cyl No = 39
Exp Date = 09/03/2011
County = 18 Oper No. = 120011



Operator Signature
WILLIAM R. PETTY

Remarks:

Form 106-I8000

80-002667
P. 99 of 99