



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>William R. Petty</u>		
Location: <u>GF AF</u>		
Intoxilyzer® Serial Number: <u>80-002667</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>39</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.08</u> %		
Test 2: <u>0.08</u> %		
Test 3: <u>0.08</u> %		
(ATTACH TEST RECORD)		

Will Petty
Chemical Test Operator Signature

13 Oct 10
Date

Deb Kashur
Reviewed By

22 Mar 2011
Date

80-002667
P. 80 of 81
RM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = GFAF 8164.13.00 06/09
10/13/2010 10:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Gas	0.080	10:36
03 Room Air	0.000	10:36
04 Std. Gas	0.080	10:37
05 Room Air	0.000	10:37
06 Std. Gas	0.080	10:38
07 Room Air	0.000	10:38

Lot No = 659358
Cyl No = 39
Exp Date = 09/03/2011
County = 18 Oper No. = 120011



Operator Signature
WILLIAM R. PETTY

Remarks:

Form 106-I8000

80-002667
P. 81 of 81
Ym