



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Chris Pulver</u>		
Location: <u>Moun trail 50</u>		
Intoxilyzer® Serial Number: <u>80-004943</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>34</u>	Expiration Date: <u>12-23-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Chris Pulver

Chemical Test Operator Signature

10-14-10
Date

Deb Kashner

Reviewed By

18 Oct 2010
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
10/14/2010 22:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:09
02 Std. Gas	0.079	22:09
03 Room Air	0.000	22:09
04 Std. Gas	0.079	22:10
05 Room Air	0.000	22:10
06 Std. Gas	0.080	22:11
07 Room Air	0.000	22:11

Lot No = 709352
Cyl No = 34
Exp Date = 12/23/2011
County = 31 Oper No. = 049912



Operator Signature
CHRISTOPHER PULVER

Remarks:

Form 106-I8000

80-004943
P. 58 of 58
scm