



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Sola Brandon</i>		
Location <i>Ward County Jail</i>	Intoxilyzer® Serial Number <i>80-004935</i>	
Gas Lot Number <i>14214080A6</i>	Gas Cylinder Number <i>70</i>	Gas Expiration Date <i>7/5/2016</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.079 AC
 B. 0.079 AC
 C. 0.079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Brandon Sola</i>	Date <i>10/16/2014</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>21 Oct 14</i>

1st

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
10/16/2014 16:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:08
02 Std. Gas	0.079	16:08
03 Room Air	0.000	16:08
04 Std. Gas	0.079	16:09
05 Room Air	0.000	16:09
06 Std. Gas	0.079	16:10
07 Room Air	0.000	16:10

Lot No = 14214080A6
Cyl No = 70
Exp Date = 07/05/2016
County = ~~26~~ 51 ~~022~~ Oper No. = 130744



Operator Signature
BRANDON SOLA

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
10/16/2014 16:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:22
02 Std. Gas	0.078	16:23
03 Room Air	0.000	16:23
04 Std. Gas	0.079	16:24
05 Room Air	0.000	16:24
06 Std. Gas	0.080	16:25
07 Room Air	0.000	16:25

Lot No = 14214080A6
Cyl No = 70
Exp Date = 07/05/2016
County = 51 Oper No. = 130744



Operator Signature
BRANDON SOLA

Remarks:

Form 106-I8000