



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Richard D. Johnson				
Location:	McLean LEC, Washburn ND.				
Intoxilyzer® Serial Number:	80-004196				
Gas Lot Number:	659357	Gas Cylinder Number:	47	Expiration Date:	9/3/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.081%					
Test 2: 0.081%					
Test 3: 0.081%					
(ATTACH TEST RECORD)					

Richard D. Johnson
Chemical Test Operator Signature

10-18-10
Date

Deb Kashner
Reviewed By

21 Oct 2010
Date

80-004196
P. 95 of 96
van


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = WASH 8164.13.00 06/09
10/18/2010 14:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:17
02 Std. Gas	0.081	14:17
03 Room Air	0.000	14:18
04 Std. Gas	0.081	14:18
05 Room Air	0.000	14:19
06 Std. Gas	0.081	14:19
07 Room Air	0.000	14:20

Lot No = 659357
Cyl No = 47
Exp Date = 09/03/2011
County = 28 Oper No. = 118707


Operator Signature
RICHARD D. JOHNSON

Remarks:

Form 106-18000

80-004196
P. 96-196
AN