

✓ CML
10-27-09

**INTOXILYZER® 5000 G AND 8000
INSTALLATION AND REPAIR CHECKOUT**

Serial No.: 80-004203 Instrument Location: Stutsman Co.

Reason for Install/Repair: _____ Install After Lab Repair/Inspection
X Other (specify) New Install

Check When Done:

- X A. Surge Protector Installed/Properly Grounded.
- X B. Telephone Line Connected to Intoxilyzer® KB.
- X C. Breath Tube Heated.
- X D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- X E. Ethanol Breath Standard: _____ Connected X Installed.
(Include Form 105-G if New Cylinder Installed.)
- X F. Review/Enter Gas Setup (Level 1, Function S). (Intoxilyzer® 8000 Only.)
- X G. Print Test (Level 1, Function P). (Attach Test Record.)
- X H. ACA Test (Level 1, Function C). (Attach Test Record.)
- X I. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- J. Repair and/or Maintenance Performed: New Install

10-21-09 AR
10-09-09
Date
26 Oct 2009
Date

[Signature]
Field Inspector's Signature
Deb Kashner
Reviewed By


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004203
Location = JAME 8164.12.00 06/09
10/21/2009 17:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:21
02 Std. Gas	0.084	17:21
03 Room Air	0.000	17:22
04 Std. Gas	0.083	17:22
05 Room Air	0.000	17:23
06 Std. Gas	0.083	17:23
07 Room Air	0.000	17:24

Lot No = 659358
Cyl No = 48
Exp Date = 09/03/2011
County = 47 Oper No. = 100414


Operator Signature
JOSHUA J. RUDE

Remarks:

Form 106-I8000

80-004203
P. J. J. Rude


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004203
Location = JAME 8164.12.00 06/09
10/08/2009 16:52

Test	AC	Time
01 Diagnostic	OK	16:53
02 Room Air	RFI*	16:54
03 Room Air	RFI*	16:54

*Invalid Test
Inhibited - RFI

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.


Operator Signature
JOSHUA J. RUDE

Remarks:

Form 106-I8000

PO-004203
P-4 of 8 PM

19 Oct 09

Hi Josh -

We need a total of 2 ACA's run for these installs. You have the ~~install~~^{ACA} for the Cylinder Install but we still need another ACA for the 8000 Install/Repair Checkout.

So please run another ACA then write the new date of this ACA on the Install/Repair sheet by crossing off the old ^{date} w/a line through it and ~~re~~ writing the new date above it, then initial. No need to change the date anywhere else.

I'm sending back all paperwork just to keep them all together. Thanks!

Call me or Margy if you have any questions.

Deb Kashur
328-6159

80-004203
P. 5 of 8

INTOXILYZER® 5000 G AND 8000
INSTALLATION AND REPAIR CHECKOUT

Serial No.: 80-004203 Instrument Location: Stutsman Co.

Reason for Install/Repair: _____ Install After Lab Repair/Inspection
X Other (specify) New Install

Check When Done:

- X A. Surge Protector Installed/Properly Grounded.
X B. Telephone Line Connected to Intoxilyzer® KB.
X C. Breath Tube Heated.
X D. Review/Enter Preliminary Data Entry (Level 2, Function E).
X E. Ethanol Breath Standard: _____ Connected X Installed.
(Include Form 105-G if New Cylinder Installed.)
X F. Review/Enter Gas Setup (Level 1, Function S). (Intoxilyzer® 8000 Only.)
X G. Print Test (Level 1, Function P). (Attach Test Record.)
X H. ACA Test (Level 1, Function C). (Attach Test Record.)
X I. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
J. Repair and/or Maintenance Performed: New Install

10-08-09
Date

[Signature]
Field Inspector's Signature

Date

Reviewed By

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004203
Location = JAME 8164.12.00 06/09
10/08/2009 16:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? Yes
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 48
Standard Expiration? 09/03/2011
Oper No? 100414

Flow Cal. Date: 06/04/2009
Slope 667
Intercept -500345

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100): -9961 -17350
1st Coef(*100): 265308 134622
2nd Coef(*100): 2401 1045
H2O adj(mg/l*10k): 485 581

***** Printer Test End *****


Operator Signature
JOSHUA J. RUDE

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004203
Location = JAME 8164.12.00 06/09
10/08/2009 16:52

Test	AC	Time
01 Diagnostic	OK	16:53
02 Room Air	RFI*	16:54
03 Room Air	RFI*	16:54

*Invalid Test
Inhibited - RFI

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.


Operator Signature
JOSHUA J. RUDE

Remarks:

Form 106-I8000

80-004203
P. 8 of 8