



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59281 (1-2010)

Serial Number: 80-003069 Instrument Location: MobL.

Reason for Install/Repair:

Install After Lab Repair/Inspection

Other (Specify) _____

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: [Signature]

[Signature]
Field Inspector Signature

10-21-10
Date

Deb Kashur
Reviewed By

11 Jan 2011
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = MOHL 8164.13.00 06/09
10/21/2010 12:43

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 27
Standard Expiration? 09/03/2011
Oper No? 039909

Flow Cal. Date: 04/08/2008
Slope 691
Intercept -599264

IR Calibration Date: 06/04/2008
3um 9um

0th Coef(*100): -14879 -13696
1st Coef(*100): 280678 136270
2nd Coef(*100): 177 615
H2O adj(mg/l*10k): 391 185

***** Printer Test End *****



Operator Signature
BARRY VANNATTA

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = MOHL 8164.13.00 06/09
10/21/2010 12:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:46
02 Std. Gas	0.079	12:46
03 Room Air	0.000	12:47
04 Std. Gas	0.079	12:47
05 Room Air	0.000	12:48
06 Std. Gas	0.080	12:48
07 Room Air	0.000	12:49

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909


Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-18000

80-003069
P. 87 of 88
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = MOHL 8164.13.00 06/09
10/21/2010 12:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	12:53
02 Room Air	0.000	12:54

*Invalid Test
Inhibited - RFI

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909



Operator Signature
BARRY VANNATTA

Remarks:

Form 106-I8000

80-003069
P. 88 of 88
JFM