



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
**OFFICE OF ATTORNEY GENERAL**  
**CRIME LABORATORY DIVISION**  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Darcy W Aberle</u>		
Location: <u>Williston LEC</u>		
Intoxilyzer® Serial Number: <u>80-004189</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>45</u>	Expiration Date: <u>09/03/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.077%</u> Test 2: <u>0.078%</u> Test 3: <u>0.077%</u>		
(ATTACH TEST RECORD)		

Darcy W Aberle  
 Chemical Test Operator Signature

10/22/2010  
 Date

Deb Kashner  
 Reviewed By

03 Mar 2011  
 Date

80-004189  
 P-110 of 111  
 PK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = WILL      8164.13.00 06/09  
10/22/2010      13:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Gas	0.077	13:39
03 Room Air	0.000	13:39
04 Std. Gas	0.078	13:40
05 Room Air	0.000	13:40
06 Std. Gas	0.077	13:41
07 Room Air	0.000	13:41

Lot No = 659357  
Cyl No = 45  
Exp Date = 09/03/2011  
County = 53      Oper No. = 010101

Darcy W Aberle  
Operator Signature  
DARCY W. ABERLE

Remarks:

Form 106-I8000

80-004189  
P. 111 of 111  
km