



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Danielle M. Roseware</u>		
Location: <u>CARR</u>		
Intoxilyzer® Serial Number: <u>80-004944</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>12</u>	Expiration Date: <u>12/03/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>.079</u> %		
Test 2: 0. <u>.079</u> %		
Test 3: 0. <u>.079</u> %		
(ATTACH TEST RECORD)		

Danielle Roseware
Chemical Test Operator Signature

10/25/10
Date

Deb Kashur
Reviewed By

~~26~~²⁷ Oct 2010
Date

80-004944
p. 55 of 56
RM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = CARR 8164.13.00 06/09
10/25/2010 14:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:02
02 Std. Gas	0.079	14:03
03 Room Air	0.000	14:03
04 Std. Gas	0.079	14:04
05 Room Air	0.000	14:04
06 Std. Gas	0.079	14:04
07 Room Air	0.000	14:05

Lot No = 709352
Cyl No = 12
Exp Date = 12/23/2011
County = 16 Oper No. = 100754

Danielle M. Roseware

Operator Signature
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004944
P. 56 of 56
DM