



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Matthew Anderson

Location: Richland County Jail

Intoxilyzer Serial No.: 80-004198

Lot Number: 659358

Cylinder Number: 52

Expiration Date: 9/30/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.081 %

Test 2: 0.081 %

Test 3: 0.081 %

(ATTACH TEST RECORD)

Matthew Anderson

Chemical Test Operator Signature

Deb Kashner

10-26-2009

Date Cylinder Changed

29 Oct 2009

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

P.1 of 2
 ym

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = WAHP 8164.12.00 06/09
10/26/2009 23:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:14
02 Std. Gas	0.081	23:14
03 Room Air	0.000	23:15
04 Std. Gas	0.081	23:15
05 Room Air	0.000	23:16
06 Std. Gas	0.081	23:16
07 Room Air	0.000	23:16

Lot No = 659358

Cyl No = 52

Exp Date = 09/03/2011

County = 39

Oper No. = 020401



Operator Signature
MATTHEW ANDERSON

Remarks:

Cylinder Installation

Form 106-I8000

*P. 2 of 2
JFM*