



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David Feland</u>		
Location: <u>Kidder S.D.</u>		
Intoxilyzer® Serial Number: <u>80-004186</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>37</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.079</u> %		
Test 2: <u>0.081</u> %		
Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

David Feland
Chemical Test Operator Signature

26 Oct 10
Date

Deb Kashner
Reviewed By

27 Oct 2010
Date

80-004186
P. 91 of 92
RM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
10/26/2010 13:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:56
02 Std. Gas	0.079	13:57
03 Room Air	0.000	13:57
04 Std. Gas	0.081	13:57
05 Room Air	0.000	13:58
06 Std. Gas	0.080	13:58
07 Room Air	0.000	13:59

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 22 Oper No. = 119404



Operator Signature
DAVID R. FELAND

Remarks:

Form 106-I8000

80-004186
P. 92 of 92
JFM