

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Pat Renz

Location: BSPD

Intoxilyzer Serial No.: 80-004186

Lot Number: 659358

Cylinder Number: 13

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 083 %

Test 3: 0. 083 %

(ATTACH TEST RECORD)

Pat Renz
Chemical Test Operator Signature

10-29-09
Date

Deb Kashur (ACA check)
Reviewed By

08 Jan 2010
Date

80-004186
P.1 of 2 PR

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = BSPD 8164.12.00 06/09
10/29/2009 12:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:03
02 Std. Gas	0.081	12:03
03 Room Air	0.000	12:04
04 Std. Gas	0.083	12:04
05 Room Air	0.000	12:05
06 Std. Gas	0.083	12:05
07 Room Air	0.000	12:06

Lot No = 659358
Cyl No = 13
Exp Date = 09/03/2011
County = 08 Oper No. = 029710

Pat Renz

Operator Signature
PATRICK M. RENZ

Remarks:

ACA

Form 106-18000

*80-004186
P. 2 of 2 KM*