



OFFICE OF ATTORNEY GENERAL
Crime Laboratory Division
2641 East Main Avenue
P.O. Box 937
Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: MITCHELL J. RUMPLE

Location: CASS County TAFE1 FARGO ND.

Intoxilyzer Serial No.: 80-002669

Lot Number: 659357

Cylinder Number: 40

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.081 %

Test 2: 0.082 %

Test 3: 0.081 %

(ATTACH TEST RECORD)

[Signature]
Chemical Test Operator Signature

11-1-09
Date Cylinder Changed

Deb Kasher
Reviewed By

15 Dec. 09
Date

Form 105-G
Toxicology Laboratory
Crime Laboratory Division

[Type text]

80-002669
P. 1 of 2 gm


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CASS 8164.12.00 06/09
11/01/2009 01:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:32
02 Std. Gas	0.081	01:32
03 Room Air	0.000	01:33
04 Std. Gas	0.082	01:33
05 Room Air	0.000	01:34
06 Std. Gas	0.081	01:34
07 Room Air	0.000	01:35

Lot No = 659357
Cyl No = 40
Exp Date = 09/03/2011
County = 09 Oper No. = 068330



Operator Signature
MITCHELL RUMPLE

Remarks:

New PAGE

Form 106-18000

80-002669
P. 2 of 2 YAM