




**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	ZLOTKOWSKI, GERARD R				
Location:	ASHL				
Intoxilyzer® Serial Number:	80-004945				
Gas Lot Number:	709352	Gas Cylinder Number:	32	Expiration Date:	12/23/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.080 %					
Test 2: 0.081 %					
Test 3: 0.080 %					
(ATTACH TEST RECORD)					

  
\_\_\_\_\_  
Chemical Test Operator Signature

11/01/2010  
\_\_\_\_\_  
Date

Deb Kashur  
\_\_\_\_\_  
Reviewed By

10 Jan 2011  
\_\_\_\_\_  
Date

80-004945  
P. 51af-52  
KFM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = ASHL      8164.13.00 06/09  
11/01/2010      17:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:39
02 Std. Gas	0.080	17:39
03 Room Air	0.000	17:40
04 Std. Gas	0.081	17:40
05 Room Air	0.000	17:40
06 Std. Gas	0.080	17:41
07 Room Air	0.000	17:41

Lot No = 709352  
Cyl No = 32  
Exp Date = 12/23/2011  
County = 26      Oper No. = 100631

  
\_\_\_\_\_  
Operator Signature  
GERARD ZLOTKOWSKI

Remarks:

Form 106-I8000

80-004945  
P. 50-150  
km