



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David M Wallace</u>		
Location: <u>Dickinson ZEC</u>		
Intoxilyzer® Serial Number: <u>80 004947</u>		
Gas Lot Number: <u>154108001</u>	Gas Cylinder Number: <u>20</u>	Expiration Date: <u>7-1-12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.028</u> %		
Test 2: <u>0.028</u> %		
Test 3: <u>0.028</u> %		
(ATTACH TEST RECORD)		

David M Wallace
Chemical Test Operator Signature

11-1-10
Date

Ddo Kashur
Reviewed By

03 NOV 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.13.00 06/09
11/01/2010 08:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:53
02 Std. Gas	0.078	08:53
03 Room Air	0.000	08:53
04 Std. Gas	0.078	08:54
05 Room Air	0.000	08:54
06 Std. Gas	0.078	08:55
07 Room Air	0.000	08:55

Lot No = 15410080A1
Cyl No = 20
Exp Date = 07/01/2012
County = 45 Oper No. = 068332


Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004947
P. 54 of 54
YH