



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

*Monthly ACA run  
DK*

**ETHANOL BREATH STANDARD CYLINDER-INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David M Wallace

Location: Dickinson 211

Intoxilyzer Serial No.: 80 004184

Lot Number: 125-935-8

Cylinder Number: 2,3

Expiration Date: 9-03-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.079 %

Test 2: 0.081 %

Test 3: 0.080 %

(ATTACH TEST RECORD)

David M Wallace  
 Chemical Test Operator Signature  
Deb Keshur

11-2-09  
 Date Cylinder Changed  
13 NOV 09

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

*80-004184  
p. 1 of 2 FM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004184  
Location = DKPD      8164.12.00 06/09  
11/02/2009      10:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:47
02 Std. Gas	0.079	10:48
03 Room Air	0.000	10:48
04 Std. Gas	0.081	10:49
05 Room Air	0.000	10:49
06 Std. Gas	0.080	10:50
07 Room Air	0.000	10:50


Lot No = 659358

Cyl No = 23

Exp Date = 09/03/2011

County = 45

Oper No. = 068332



Operator Signature  
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004184  
P. 2 of 2 JH