



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Tyler Alan Wolfe</u>		
Location: <u>HARV</u>		
Intoxilyzer® Serial Number: <u>80-004935</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>12/23/2011</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>075</u> %		
Test 2: 0. <u>077</u> %		
Test 3: 0. <u>076</u> %		
(ATTACH TEST RECORD)		

Tyler Wolfe  
Chemical Test Operator Signature

11-2-10  
Date

Deb Kashur  
Reviewed By

05 NOV 2012  
Date


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004935  
Location = HARV      8164.13.00 06/09  
11/02/2010      15:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:46
02 Std. Gas	0.075	15:47
03 Room Air	0.000	15:47
04 Std. Gas	0.077	15:48
05 Room Air	0.000	15:48
06 Std. Gas	0.076	15:48
07 Room Air	0.000	15:49

Lot No = 709352  
Cyl No = 31  
Exp Date = 12/23/2011  
County = 52      Oper No. = 060901

  
\_\_\_\_\_  
Operator Signature  
TYLER A. WOLFE

Remarks:

ACA

Form 106-I8000

80-004935  
P. 63 of 63  
van