



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Daniel Haugen

Location: Rolla

Intoxilyzer Serial No.: 80-004199

Lot Number: 659358

Cylinder Number: 11

Expiration Date: 9-03-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 080 %

Test 3: 0. .080 %

(ATTACH TEST RECORD)

Daniel Haugen
 Chemical Test Operator Signature

11-03-09
 Date Cylinder Changed

Deb Kashner
 Reviewed By

17 NOV 09
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004199
 P. 1 of 2 PH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = ROLL 8164.12.00 06/09
11/03/2009 15:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:59
02 Std. Gas	0.082	15:59
03 Room Air	0.000	16:00
04 Std. Gas	0.080	16:00
05 Room Air	0.000	16:01
06 Std. Gas	0.080	16:01
07 Room Air	0.000	16:02

Lot No = 659358

Cyl No = 11

Exp Date = 09/03/2011

County = 40

Oper No. = 010105



Operator Signature
DANIEL J. HAUGEN

Remarks:

*cylinder install
Form 105-6*

Form 106-I8000

*80-004199
P. 2 of 2 fm*