



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>H. Tracy L. Meidinger #258</u>		
Location: <u>UND-PD</u>		
Intoxilyzer® Serial Number: <u>80-004934</u>		
Gas Lot Number: <u>654358</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>09-30-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>081</u> %		
Test 2: 0. <u>081</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

H. Tracy L. Meidinger #258  
 Chemical Test Operator Signature

11-03-2010  
 Date

Deb Kashner  
 Reviewed By

08 NOV 2010  
 Date

80-004934  
 P. 59 of 60  
 MEX

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = UNDP      8164.13.00 06/09  
11/03/2010      21:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:09
02 Std. Gas	0.081	21:09
03 Room Air	0.000	21:09
04 Std. Gas	0.081	21:10
05 Room Air	0.000	21:10
06 Std. Gas	0.080	21:11
07 Room Air	0.000	21:11

Lot No = 659358  
Cyl No = 31  
Exp Date = 09/03/2011  
County = 18      Oper No. = 119409

*Tracy L. Meidinger #258*  
Operator Signature  
TRACY L. MEIDINGER

Remarks:

Form 106-I8000

ACA

80-004934  
P. 20 of 60  
vsm