



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <i>Matthew Anderson</i>		
Location: <i>Richland County Jail</i>		
Intoxilyzer® Serial Number: <i>80-004206</i>		
Gas Lot Number: <i>659358</i>	Gas Cylinder Number: <i>52</i>	Expiration Date: <i>9-03-11</i>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <i>0.083</i> % Test 2: <i>0.083</i> % Test 3: <i>0.084</i> %		
(ATTACH TEST RECORD)		

*Matthew Anderson*  
 \_\_\_\_\_  
 Chemical Test Operator Signature

*11-04-10*  
 \_\_\_\_\_  
 Date

*Deb Kashner*  
 \_\_\_\_\_  
 Reviewed By

*08 NOV 2010*  
 \_\_\_\_\_  
 Date

*80-004206*  
*P. 108 of 109*  
*SM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004206  
Location = WAHP      8164.13.00 06/09  
11/04/2010      15:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:02
02 Std. Gas	0.083	15:02
03 Room Air	0.000	15:03
04 Std. Gas	0.083	15:03
05 Room Air	0.000	15:03
06 Std. Gas	0.084	15:04
07 Room Air	0.000	15:04

Lot No = 659358  
Cyl No = 52  
Exp Date = 09/03/2011  
County = 39      Oper No. = 020401



Operator Signature  
MATTHEW W. ANDERSON

Remarks:

Form 106-I8000

80-004206  
P. 109 of 148  
JAN