



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Steven R Föcher</u>		
Location: <u>Rolette County Sheriff's Office, Rolla, ND</u>		
Intoxilyzer® Serial Number: <u>80-004199</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>11</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>081</u> %		
Test 2: 0. <u>081</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

11-6-10
Date

Deb Kashur
Reviewed By

16 NOV 10
Date

80-004199
p. 93 of 94
jam

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = ROLL 8164.13.00 06/09
11/06/2010 20:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:51
02 Std. Gas	0.081	20:51
03 Room Air	0.000	20:52
04 Std. Gas	0.081	20:52
05 Room Air	0.000	20:52
06 Std. Gas	0.080	20:53
07 Room Air	0.000	20:53


Lot No = 659358

Cyl No = 11

Exp Date = 09/03/2011

County = 40

Oper No. = 049904



Operator Signature
STEVEN R. FISCHER

Remarks:

Form 106-I8000

80-004199
p. 94 of 94
JFM