



OFFICE OF ATTORNEY GENERAL  
Crime Laboratory Division  
2641 East Main Avenue  
P.O. Box 937  
Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Officer Pete Nielsen

Location: West Fargo PD

Intoxilyzer Serial No.: 80-003065

Lot Number: 659357

Cylinder Number: 10

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 080 %

Test 2: 0. 080 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)

Chemical Test Operator Signature

11-13-2009  
Date Cylinder Changed

Deb Kashner  
Reviewed By

19 NOV 2009  
Date

Form 105-G  
Toxicology Laboratory  
Crime Laboratory Division

[Type text]

80-003065  
P. 1 of 2 FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = WFPD      8164.12.00 06/09  
11/13/2009      12:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:40
02 Std. Gas	0.080	12:40
03 Room Air	0.000	12:41
04 Std. Gas	0.080	12:41
05 Room Air	0.000	12:42
06 Std. Gas	0.081	12:42
07 Room Air	0.000	12:43


Lot No = 659357

Cyl No = 10

Exp Date = 09/03/2011

County = 09

Oper No. = 039617



Operator Signature  
PETER D. NIELSEN

Remarks:

Form 106-I8000

80-003065  
P. 2 of 2 FM